

DECLARATION OF CONFORMITY / DÉCLARATION DE CONFORMITÉ

Application of Council Directives / <i>Application des directives:</i>	2014/30/EU - EMC Directive 2014/35/EU – Low Voltage Directive 2011/65/EU – RoHS Directive
Standards to which conformity is declared / <i>Conformité selon les normes suivantes:</i>	IEC 61010-1-010:2010 3rd ed. IEC 61010-2-010: 2014 3rd ed. IEC 61010-2-081: 2015 2nd ed. EN61326 : 2013 IEC 61000-4-2: 2008 IEC 61000-4-3: 2006, +A1: 2007, +A2: 2010 IEC 61000-4-4:2012 IEC 61000-4-5 : 2005 ed2.0 (2005-11) IEC 61000-4-6: 2008 IEC 61000-4-8: 2009 IEC 61000-4-11: 2004 CISPR 11: 2009 A1: 2010, Group 1, Class A FCC Part 15: 2012, Subpart B, Class A EN50581:2012
Name of Manufacturer / <i>Nom du fabricant:</i>	SCP SCIENCE
Manufacturer's address / <i>Adresse du fabricant:</i>	21800 Clark Graham Baie D'Urfé, Quebec Canada, H9X 4B6
Type of equipment / <i>Type d'équipement:</i>	EasyPREP
Model / Modèle:	COD-200, BOD-200, BOD-300, Sample Handler, Titrec
Code:	TBD

I, the undersigned, hereby declare that the equipment specified above conforms to the above Standards /
Je soussigné déclare que l'équipement spécifié sur ce certificat est conforme aux normes indiquées ci-haut:

Name / <i>Nom:</i>	David Smith
Title / <i>Titre:</i>	Quality Assurance Manager / Directeur Assurance Qualité
Signature:	
Date (mm/dd/yy):	08/15/2019

**CORPORATE
HEADQUARTERS**

21800 Clark Graham
Baie D'Urle (Montréal), Québec
H9X 4B6, Canada
Phone: +1 (800) 361-6820
Fax: +1 (800) 253-5549

USA

P.O. Box 3221
Champlain, NY,
12919-4816 USA
Phone: +1 (800) 361-6820
Fax: +1 (800) 253-5549

FRANCE

12 Ave. de Québec, Bat. Iberis
SILC 642, 91965 Courtaboeuf,
France
Phone: +33 (0) 1 69 18 71 17
Fax: +33 (0) 1 60 92 05 67

GERMANY

Alte Marktoberdorfer Straße 14,
87616 Marktoberdorf
Phone: +49 (0) 8342-89560-61
Fax: +49 (0) 8342-89560-69

CORPORATE:

Phone: +1 (514) 457-0701 | Fax: +1 (514) 457-4499

www.scpscience.com | sales@scpscience.com