

CE Declaration of Conformity

Application of Council Directive(s):

EN 61010-1: Safety requirements for electrical equipment for measurement, control and laboratory use
- Part 1: General requirements (IEC 61010-1:2001);

EN 61326): Electrical equipment for measurement, control and laboratory use - EMC requirements
(IEC 61326-1:1997 + IEC 61326-1/A1:1998 + IEC 61326-1/A2:2000 + Annexes E & F of IEC
61326:2002 + Corrigendum:2002);

| Test | Standard |
|---|-----------------|
| Radiated E-field emissions, 30-1000 MHz | EN 61326-1:2006 |
| Conducted emissions, 0.15 - 30 MHz | EN 61326-1:2006 |
| Electrostatic discharge (ESD) | EN 61000-4-2 |
| Radiated RF immunity | EN 61000-4-3 |
| Electrical fast transient/burst | EN 61000-4-4 |
| Surge immunity | EN 61000-4-5 |
| Conducted RF immunity | EN 61000-4-6 |
| Voltage dips and interruptions | EN 61000-4-11 |

| | |
|-------------------------------|--|
| Manufacturer's Name | inXitu inc. |
| Manufacturer's Address | 2551 Casey Ave, Suite A Mountain View, CA 94043 |
| Type of Equipment | Scientific, Test, Measurement equipment |
| Models | Terra, BTX |
| Serial Number | N/A |

*I, the undersigned, hereby declare that the equipment specified above
conforms to the above Directive(s) and Standard(s)*

Place: Mountain View, California USA

Date: October 25, 2009



Bradley W Boyer
President/CEO inXitu inc.