

Application Note - Digestion of Apple-Lemon Flavoured Powder

Introduction:

This study evaluated the effectiveness of the *NOVAWAVE* microwave digestion system to digest food samples.

Sample Type:

Apple-Lemon flavoured powder

- Sample weight: 0.25-0.50 g
- 3 replicates

Supplies and Reagent:

- 1) Microwave Digestion System, *NOVAWAVE* SA Model*
- 2) Quartz Vessels, 75 ml, calibrated at 50ml*
- 3) Teflon® Caps and Safety Pressure Release Caps (pre-set release pressure 435psi) (or 30 bar)*
- 4) Nitric Acid, *Plasma*PURE HNO₃ (70%)*
- 5) ICP-MS Spectrometer, ELAN 6100, Perkin-Elmer
- 6) Nebulizer*
- 7) Alumina Injector Torch, 1.2mm*
- 8) Cyclonic Spray Chamber, Baffled*
- 9) Graduated Cylinder, Corning
- 10) Volumetric flask, 100 ml

* Manufactured by SCP SCIENCE

Sample Preparation Procedure:

The samples were weighed on a 4 place analytical balance directly in the 75 quartz vessels. After adding 10ml of HNO₃ and the appropriate spike, the samples were allowed to sit at room temperature for 5-10 minutes. The vessels were placed in the rack, capped and digested following the *NOVAWAVE* heating profile. At the end of the digestion, the solutions were clear and no particulates remain. After cooling to room temperature, the rack was vented carefully in a fumehood. The samples were normalized to 100 ml with deionized water in a volumetric flask and analyzed on the ICP-MS.

SCP SCIENCE

Providing Innovative Solutions to Analytical Chemists.....

www.scpscience.com
sales@scpscience.com

A new category of Automated Microwave Digestion

Application Note - Digestion of Apple-Lemon Flavoured Powder

NOVAWAVE Heating Program :

STAGE	RAMP TIME (Minutes)	PRESSURE (PSI – LIMIT)	TEMPERATURE °C	HOLD TIME (Minutes)
1	20	435	200	15

Results:

Apple-Lemon flavoured powder

El.	AMU	NOVAWAVE (PPM)	SPIKE (PPM)	NOVAWAVE SPIKE SAMPLE 1 (PPM)	NOVAWAVE SPIKE SAMPLE 2 (PPM)	NOVAWAVE SPIKE SAMPLE AVG (PPM)	RECOVERY (%)
As	75	0.0011	0.025	0.036	0.044	0.040	116
Cd	111	0.003	0.5	0.520	0.522	0.521	104
Hg	202	0.023	0.025	0.049	0.055	0.052	116
Pb	206	0.763	0.5	1.307	1.295	1.301	108

Conclusion

The *NOVAWAVE* affords good results according to the spiked standard addition method. All recovery are within 80-120% of the spiked value.

SCP SCIENCE

Providing Innovative Solutions to Analytical Chemists

CANADA / USA

Tel.: (800) 361-6820

Fax: (800) 253-5549

FRANCE

Tel.: +33 (0) 1 69 18 71 17

Fax: +33 (0) 1 60 92 05 67

INTERNATIONAL

Tel.: (514) 457-0701

Fax: (514) 457-4499

CHINA

Tel.: +86 (10) 58032301

Fax.: +86 (10) 58032302

BRAZIL

Tel: +55 19 8194-8488

www.scpscience.com
sales@scpscience.com